Draft Medical device product items that shall Accord specification, testing method and performance

Classification	Classification	Class	Product	Specification, Testing Method and Performance	Implementation
Number	Name		Category		Date
O.3825	Mechanical	1	Walking	According to Product Characteristics, Non-clinical	One year after
	Walker		Frames	performance data of the Fatigue test, Static loading	announcement
			without	test, Static leg-strength test, Forward stability test	
			Accessories	and Backward stability test should comply with CNS	
				15037-1, ISO 11199-1 or the others equivalent	
				international standards.	
			Rollators	According to Product Characteristics, Non-clinical	
				performance data of the Forward-direction stability	
				test, Backward-direction stability test, Sideway-	
				direction stability test, Brake tests, Handgrip test,	
				Rubber tip test, Resting seat test, Static loading test,	
				Fatigue test and Final inspection should comply with	
				CNS 15037-2, ISO 11199-2 or the others equivalent	
				international standards.	
			Walking	According to Product Characteristics, Non-clinical	
			Tables	performance data of the Forward stability test,	
				Backward stability test, Sideways stability test,	
				Brake test, Resting seat test, Static strength test,	
				Fatigue test and Final inspection should comply with	
				CNS 15037-3, ISO 11199-3 or the others equivalent	
				international standards.	
G.3300	Hearing Aid	1	Air-	According to Product Characteristics, Non-clinical	One year after
			conduction	performance data should comply with IEC 60601-	announcement
			Hearing Aid	1(Electrical safety test), IEC 60601-1-	
				2(Electromagnetic compatibility test), ISO 10993-	
				1(Cytotoxicity), ISO 10993-5(Sensitization), ISO	
				10993-10(Irritation or Intracutaneous reactivity) and	
				ANSI S3.22(Electroacoustic performance) or the	
				others equivalent international standards.	
I.4040	Medical	1	Disposable	According to Product Characteristics, Non-clinical	One year after
	Apparel		Isolation	performance data of the Hydrostatic pressure and	announcement
			Gown	Impact penetration should comply with CNS	
				14798(T5019) P1 level or the others equivalent	
				international standards.	